



## Arthroscopic Electrodes

Single Patient Use Only  
Disposable • Sterile  
**Rx Only**



### IMPORTANT INFORMATION

Read carefully and completely before use.

### INTENDED USE

DeRoyal Arthroscopic Electrodes are intended to be used for cutting and coagulation of soft tissue. The disposable Arthroscopic Electrodes are available in a variety of tip configurations. The electrodes are designed to fit most standard electrosurgical pencils that accommodate an electrode with a 3/32"/2.4mm shaft.

### INDICATIONS

DeRoyal Arthroscopic Electrodes are specifically designed for use in arthroscopic surgery where monopolar electrosurgical cutting and coagulation are desired.

### CONTRAINDICATIONS

DeRoyal Arthroscopic Electrodes are NOT to be used in non-arthroscopic surgical procedures.

### DIRECTIONS FOR USE

1. When inserting and/or changing the electrode, verify that the pencil is not connected to the generator or that the generator is off (in stand-by).
2. Insert the electrode into the pencil. Prior to use, check the electrode connection

with the pencil to ensure proper fit and compatibility.

3. Remove the tip protector.
4. Grasp the end of the electrode.
5. Ensure that the cap fits securely on the nose of the pencil. Improper installation may result in injury to the patient or operating room personnel by arcing at the electrode/pencil connection.
6. Ensure that conductive tissue or other objects are not near the pencil/electrode connection when the pencil is activated.

### WARNINGS

If end cap does not fit securely over the nose to the electrode, do not use the electrode. The patient or surgical personnel may be shocked or burned.

**Contents sterile unless package has been opened or damaged.**



**Manufacturer**  
DeRoyal  
200 DeBusk Lane, Powell, TN 37849 U.S.A.  
888.938.7828

©2009 DeRoyal. All Rights Reserved.  
Part # 4-10488 Revised 12/09



## Arthroscopic Electrodes

Single Patient Use Only  
Disposable • Sterile  
**Rx Only**



### IMPORTANT INFORMATION

Read carefully and completely before use.

### INTENDED USE

DeRoyal Arthroscopic Electrodes are intended to be used for cutting and coagulation of soft tissue. The disposable Arthroscopic Electrodes are available in a variety of tip configurations. The electrodes are designed to fit most standard electrosurgical pencils that accommodate an electrode with a 3/32"/2.4mm shaft.

### INDICATIONS

DeRoyal Arthroscopic Electrodes are specifically designed for use in arthroscopic surgery where monopolar electrosurgical cutting and coagulation are desired.

### CONTRAINDICATIONS

DeRoyal Arthroscopic Electrodes are NOT to be used in non-arthroscopic surgical procedures.

### DIRECTIONS FOR USE

1. When inserting and/or changing the electrode, verify that the pencil is not connected to the generator or that the generator is off (in stand-by).
2. Insert the electrode into the pencil. Prior to use, check the electrode connection

with the pencil to ensure proper fit and compatibility.

3. Remove the tip protector.
4. Grasp the end of the electrode.
5. Ensure that the cap fits securely on the nose of the pencil. Improper installation may result in injury to the patient or operating room personnel by arcing at the electrode/pencil connection.
6. Ensure that conductive tissue or other objects are not near the pencil/electrode connection when the pencil is activated.

### WARNINGS

If end cap does not fit securely over the nose to the electrode, do not use the electrode. The patient or surgical personnel may be shocked or burned.

**Contents sterile unless package has been opened or damaged.**



**Manufacturer**  
DeRoyal  
200 DeBusk Lane, Powell, TN 37849 U.S.A.  
888.938.7828

©2009 DeRoyal. All Rights Reserved.  
Part # 4-10488 Revised 12/09